



Ministry of Health and Wellness

MHPQ/EQ/2023-2024/Q26

To all Potential Bidders

**ADDENDUM NO. 1**

**Supply, Installation and Commissioning of Lithotripsy Machine for Victoria Hospital**

Please refer to our invitation for Bids dated **18 December 2023** on the above subject.

2. Following request for clarifications from potential bidders, please find below the corresponding information:

Item No	Clarification Requested	Reply
Item No 1 Line 1.6	We supply real time ultrasound localization system. Kindly confirm whether real time echography machine will be acceptable for the tender?	The equipment should have both x-ray and ultrasound for the localization of urinary stone. Please note that in line ultrasound will be used for localization and therapeutic purposes. However, an alternative method providing the facility for localization and therapeutic purpose using both x-ray and ultrasound having minimal deflection effects will be considered. Bidders are requested to submit all relevant documentary evidence pertaining to the technology used for accurate localization and visualization of stones including radiolucent stone.
Item No 1 Line 7.2 (2)	We supply two heads of 1,000,000 each. Since 2 million shots are for specific brands, kindly confirm whether 1 head with 1 million shots should be provided post warranty period or 2 heads with 1 million shots each should be provided during post warranty period?	Line specification 7.2(2) is being maintained, the new shock wave head should bear a full warranty for 2 million shots.

Item No 1 Line 3.1(m)	We supply two heads of 1,000,000 counts each. Since 2 million shots is for specific brands, kindly confirm whether: 1. 1 head with 1 million shots can be supplied, or 2. 2 heads with 1 million shots each can be supplied	Line specification 3.1(m) of <b>at least 2</b> million shots is being maintained.
Item No 1 Line 5	Please confirm the size of item No.5.1 and 5.2.	Size to be provided should be large (L).
Item No 1 Line 3.1	As we recommended using 1 million shocks per head as energy will be high and stable throughout the entire lifetime of the head providing maximum efficiency and excellent results.	Line specification 3.1(m) of <b>at least 2</b> million shots is being maintained
Item No 1 Line 1.6	Kindly note that with the inline use of ultrasound, user will not be able to do Xray simultaneously thus resulting limited functionality.	The equipment should have both x-ray and ultrasound for the localization of urinary stone. Please note that in line ultrasound will be used for localization and therapeutic purposes. However, an alternative method providing the facility for localization and therapeutic purpose using both x-ray and ultrasound having minimal deflection effects will be considered. Bidders are requested to submit all relevant documentary evidence pertaining to the technology used for accurate localization and visualization of stones including radiolucent stone.

3. Bidders are kindly requested to take note and submit their bids accordingly. This addendum should be inserted in the Bid Documents upon submission.

**Senior Chief Executive  
Ministry of Health and Wellness  
5<sup>th</sup> floor, Emmanuel Anquetil Building  
SSR Street, Port Louis  
04 January 2024**