



B7C
1378

Ministry of Health and Wellness

MHPQ/EQ/2023-2024/Q38

Addendum No. 1

Supply, Installation and Commissioning of Combined Biometer with Keratometer & other Medical Equipment for New Moka Eye Hospital – MHPQ/EQ/2023-2024/Q38

To: All Potential Bidders

The Ministry has carried out a bidding exercise on the above subject on 03 April 2024. The closing date for submission of bids is 29 April 2024.

2. Following request for clarification from potential bidders, please find below the corresponding information:

Item No.	Line Specification	Clarification Requested	Reply
Item No. 1 – Combined Biometer with Keratometer	1.28	As all optical biometers currently in the market, to our knowledge operate with a joystick for a precise maneuverability and optimal functionality, there is no requirement for a touchpad interface.	As per market survey, the equipment which is a combined system as per tender can be controlled by a joystick and computer keyboard with integrated touchpad. Alternatively, other proposals with innovative and ergonomic interface and hardware to enable smooth operability and control of the equipment will also be acceptable. Bidders must describe and provide documentary evidence of how the hardware in their equipment are designed for smooth user operation both for biometry and keratometry.
Item No. 4 – Automated Perimeter Visual Field Analyser	4.16.3	We would like to request if point 4.16.3 can be amended to "magnetic or motorized chin rest and forehead rest" which would be a more inclusive term and encourage fair competition.	The proposal for "magnetic chin rest OR motorized chin rest and forehead rest" is acceptable.

3. You are therefore requested to:

- (a) take into account the whole of addendum No. 1 whilst submitting your bid; and
- (b) enclose the whole of addendum No. 1 in your bid document while submitting your bid.

Senior Chief Executive
Ministry of Health and Wellness
5th Floor, Emmanuel Anquetil Building
SSR Street, Port Louis
19 April 2024